



CERTIFICATE No. 18/0916/00/0161

PPE TYPE FILTERING HALF MASK
REF: FMP3V FFP3 NR D



AITEX, Notified Body No. 0161 for the application of Regulation (EU) 2016/425 of the European Parliament and of the Council of 9th March 2016, in which the essential health and safety requirements that Personal Protective Equipment (PPE) must comply with.

CERTIFIES The Company:

ASATEX AG
August-Borsig-Str. 2
DE-50126 Bergheim
GERMANY



CERTIFICATE No. 18/0916/00/0161

Has obtained EU TYPE EXAMINATION in compliance with what is stated in Regulation (EU) 2016/425 and in agreement with the applicable test procedures and technical specifications

Destined for respiratory protection, according to the following standard/s:

- EN 149:2001+A1:2009. Respiratory protective devices as filtering half masks to protect against particles, except for those adapted for escape situations, class FFP3 NR D, with the optional requirement of clogging.

Having achieved the performance requirements specified in Technical Test Report No. **2018EC0053UE** and the PPE's Technical Documentation.

Description of the PPE:

The particle filtering half mask covers nose and chin.

The particle filtering half mask is composed by a principal body with a head harness, a nose clip, a nose cushion and an exhalation valve.

The PPE is made from the following materials according to the information supplied by the costumer:

- Main body of the particle filtering half mask:
 - Polyester, the inner layer.
 - Polypropylene filter fibre, the filter media.
 - Polypropylene, the outer layer.
- Head harness: cotton with lycra, free of latex.
- Strap block: nylon plastic.
- Nose clip: aluminium strip.
- Nose cushion: polyurethane foam.
- Exhalation valve: high density polyethylene diaphragm of siliconized rubber.

It shall be the manufacturer's responsibility to provide specific information of this certificate and the tested levels of protection.

The CAT. III PPE shall only be used in conjunction with one of the conformity assessment procedures according to module C2 or module D described in article 19 letter c) of the Regulation (EU) 2016/425.

Digitally Signed by: Silvia Devesa
Date: 05/12/2018 16:29:55
Location: Alcoy

Silvia Devesa Valencia
Laboratory Subdirector and Innovation

Date of issue of the Certificate: 5th of December 2018
Date of expiry: 5th of December 2023